










CRITICAL REVIEW

Feasibility of cardiac-based seizure detection and prediction: A systematic review of non-invasive wearable sensor-based studies

Eryse Amira Seth^{1,2}  | Jessica Watterson^{2,3}  | Jue Xie³  | Alina Arulsamy^{1,2}  |
 Hadri Hadi Md Yusof^{1,2}  | Irma Wati Ngadimon^{1,2}  | Ching Soong Khoo⁴  |
 Amudha Kadirvelu²  | Mohd Farooq Shaikh^{1,2,5} 

¹Neuropharmacology Research Laboratory, Jeffrey Cheah School of Medicine and Health Sciences, Monash University Malaysia, Bandar Sunway, Malaysia

²Jeffrey Cheah School of Medicine and Health Sciences, Monash University Malaysia, Bandar Sunway, Malaysia

³Department of Human-Centred Computing, Monash University, Melbourne, Victoria, Australia

⁴Neurology Unit, Department of Medicine, Universiti Kebangsaan Malaysia Medical Centre, Kuala Lumpur, Malaysia

⁵School of Dentistry and Medical Sciences, Charles Sturt University, Orange, New South Wales, Australia

Correspondence

Mohd Farooq Shaikh, School of Dentistry and Medical Sciences, Charles Sturt University, 346 Leeds Parade, Orange, NSW 2800, Australia.

Email: farooq.shaikh@monash.edu and mshaikh@csu.edu.au

Funding information

Monash University Malaysia NEED (NETwork for Equity through Digital Health), Grant/Award Number: MED/NEED/11-2020/002

Abstract

A reliable seizure detection or prediction device can potentially reduce the morbidity and mortality associated with epileptic seizures. Previous findings indicating alterations in cardiac activity during seizures suggest the usefulness of cardiac parameters for seizure detection or prediction. This study aims to examine available studies on seizure detection and prediction based on cardiac parameters using non-invasive wearable devices. The Embase, PubMed, and Scopus databases were used to systematically search according to the Preferred Reporting Items for Systematic Reviews and Meta-Analysis guidelines. Human studies that evaluated seizure detection or prediction based on cardiac parameters collected using wearable devices were included. The QUADAS-2 tool and proposed standards for validation for seizure detection devices were used for quality assessment. Twenty-four articles were identified and included in the analysis. Twenty studies evaluated seizure detection algorithms, and four studies focused on seizure prediction. Most studies used either a wrist-worn or chest-worn device for data acquisition. Among the seizure detection studies, cardiac parameters utilized for the algorithms mainly included heart rate (HR) (n = 11) or a combination of HR and heart rate variability (HRV) (n = 6). HR-based seizure detection studies collectively reported a sensitivity range of 56%-100% and a false alarm rate (FAR) of 0.02-8/h, with most studies performing retrospective validation of the algorithms. Three of the seizure prediction studies retrospectively validated multimodal algorithms, combining cardiac features with other physiological signals. Only one study prospectively validated their seizure prediction algorithm using HRV extracted from ECG data collected from a custom wearable device. These studies have demonstrated the feasibility of using cardiac parameters for seizure detection and prediction with wearable devices, with varying algorithmic performance. Many studies are in the proof-of-principle stage, and evidence for

This is an open access article under the terms of the [Creative Commons Attribution-NonCommercial-NoDerivs](https://creativecommons.org/licenses/by-nc-nd/4.0/) License, which permits use and distribution in any medium, provided the original work is properly cited, the use is non-commercial and no modifications or adaptations are made.

© 2023 The Authors. *Epilepsia Open* published by Wiley Periodicals LLC on behalf of International League Against Epilepsy.

real-time detection or prediction is currently limited. Future studies should prioritize further refinement of the algorithm performance with prospective validation using large-scale longitudinal data.

Plain Language Summary

This systematic review highlights the potential use of wearable devices, like wristbands, for detecting and predicting seizures via the measurement of heart activity. By reviewing 24 articles, it was found that most studies focused on using heart rate and changes in heart rate for seizure detection. There was a lack of studies looking at seizure prediction. The results were promising but most studies were not conducted in real-time. Therefore, more real-time studies are needed to verify the usage of heart activity-related wearable devices to detect seizures and even predict them, which will be beneficial to people with epilepsy.

KEYWORDS

cardiac, heart rate, seizure detection, seizure prediction, wearable device

1 | INTRODUCTION

Epileptic seizures are associated with an increased risk of depression, anxiety, seizure-related injuries, and premature death, known as sudden unexpected death in epilepsy (SUDEP).^{1,2} The lifetime prevalence of epilepsy, including cases in remission, is 7.60 per 1000 people overall.³ A significant number of people with epilepsy (PWE) still experience inadequate seizure control, despite considerable progress in treatment and surgical interventions.⁴ Due to the unpredictability of seizures, physicians are reliant on patients and caregivers to document seizure events.⁵ However, self-reporting is often unreliable and inaccurate,^{6–8} posing challenges to timely and effective treatment, self-management, and the risk of seizure-related injuries and SUDEP.

Recent technological advancements have paved the way for improved treatment and management strategies through seizure detection and prediction. Seizure detection is the identification of a seizure upon onset, providing objective seizure quantification,⁹ while seizure prediction involves identifying physiological changes preceding a seizure and alerting patients and caregivers of a seizure risk at any given time. An online survey conducted by the Epilepsy Innovation Institute (Ei²) revealed that unpredictability was the most hindering aspect for PWE.¹⁰ A reliable seizure prediction system may reduce anxiety, improve quality of life, and potentially eliminate the risk of injuries and SUDEP. It can also be used in the context of treatment, as medications could be titrated according to periods of high or low seizure likelihood, further improving patient adherence and side effects.^{11,12}

Research on seizure detection or prediction based on non-cerebral signals has grown significantly due to the rising prevalence of wearable devices that can non-invasively

Key Points

- There is promising evidence for seizure detection and prediction based on cardiac parameters using wearable devices.
- Most of the studies aimed to develop seizure detection algorithms, with only a few studies focusing on seizure prediction.
- Cardiac parameters used included heart rate, heart rate variability, or a combination of both, yielding diverse algorithm performance.
- Future studies should focus on prospectively validating algorithms with large-scale longitudinal data to enhance algorithm performance.

measure signals such as accelerometer (ACC), electrocardiogram (ECG), electrodermal activity (EDA), and electromyography.^{13–16} By coupling these measurements with the application of machine learning tools, substantial progress has been made in generating new insights into seizure patterns (Figure 1). Dysfunction in the autonomic nervous system (ANS) has been particularly observed in focal seizures with a temporal lobe origin, as well as focal-to-bilateral and generalized tonic-clonic seizures.¹⁷ This systematic review mainly focuses on the cardiac changes associated with epileptic seizures as an indicator of seizure onset, given the compelling evidence for pre-ictal and ictal cardiac manifestations.^{17,18} Alterations in heart rate (HR) are the most commonly observed ictal autonomic changes and could potentially serve as the earliest clinical sign of an impending seizure.¹⁷ This includes ictal tachycardia^{19,20} or a decrease in




FIGURE 1 Schematic diagram of a seizure detection/prediction system. Physiological signals collected from wearable devices undergo pre-processing, and biomarkers for seizure detection/prediction are extracted. Analysis of these biomarkers is followed by the classification step, which is used to make a decision, triggering an alert notifying users of an upcoming seizure.

heart rate variability (HRV),²¹ which is the variation in time intervals between successive heartbeats. HRV is a reflection of cardiac activity regulation by the ANS, suggesting its potential value in the identification of an upcoming seizure.²¹

In this systematic review, we aim to examine currently available studies on seizure detection or prediction based on cardiac parameters using non-invasive wearable devices and to compare the performance between different cardiac parameters.

2 | METHODS

2.1 | Search strategy

The Scopus, PubMed, and Embase databases were used to conduct a systematic search in accordance with the Preferred Reporting Items for Systematic Reviews and Meta-Analysis (PRISMA) guidelines. All available publications up to August 2022 were included, although it is worth noting that the utilization of sensor technology has only emerged in the last decade.^{22,23} Keywords related to [“epilepsy” or “seizure”] were combined with terms related to [“detection” or “prediction”], [“heart rate” or “cardiac”] and [“wearable device”]. Searches in all databases were conducted based on title and abstract.

2.2 | Study selection

All resulting articles were imported into Covidence software (Veritas Health Innovation), and duplicates were automatically removed. One reviewer screened titles and abstracts to identify relevant research articles. Two independent reviewers screened the full-text articles for inclusion, and conflicts were resolved by a third reviewer. The following inclusion criteria were used: (a) written in English; (b) observational studies (prospective or retrospective studies) involving human participants; (c) peer-reviewed original research articles related to seizure detection or prediction using wearable devices in people with epilepsy; (d) ECG or cardiovascular parameters were used as a basis for seizure detection or prediction, either alone or with other

physiological signals, (e) provided at least one algorithm performance indicator as an outcome. We also included studies that analyzed blood volume pulse (BVP) obtained from photoplethysmography (PPG) signals, as it provides information about heart rate.²⁴ Studies were excluded based on the following criteria: (a) articles identified as conference papers, reviews, book chapters, commentaries, editorials, and case reports; (b) ECG or cardiovascular data not collected using a wearable device; (c) studies that did not use cardiac parameters as a basis for seizure detection or prediction; (d) studies involving neonates.

2.3 | Data extraction and synthesis of results

Data from the included studies were extracted electronically using Covidence. Data were extracted in the following categories: study identifiers (author, year of publication), study characteristics (study population, study setting, reference standard, total participants recruited, and number of patients analyzed), wearable device (wearable device used, device location, and physiological signal[s] collected), seizure detection or prediction algorithm (type of validation, detection or prediction, modality, cardiovascular parameter[s] used), and results (seizure type[s] and algorithm performance). Studies on seizure detection were analyzed separately from those on seizure prediction, which also included seizure forecasting.

2.4 | Quality assessment

Two reviewers independently conducted the quality assessment for all included studies, and conflicts were resolved by discussion. The QUADAS-2 tool, a risk of bias tool that can be applied to primary diagnostic accuracy studies, was used to assess the quality of the studies included in the review using Review Manager version 5.4 (Cochrane Collaboration). The risk of bias was assessed based on each of these four domains: patient selection, index test, reference standard, flow, and timing. The

patient selection domain assesses the method of patient recruitment and the patients included in the study. The index test and reference standard domains assess how they were conducted and interpreted, where interpretation of the index test results may be influenced by the knowledge of the reference standard and thus introduces the potential for bias. Concerns regarding applicability to the review question were also assessed for the first three domains. The flow and timing domain assesses the inclusion of all patients in the analysis and the interval between the index test and reference standard.²⁵ All included studies were also evaluated based on proposed standards for clinical validation of seizure detection devices.²⁶ The studies were categorized into five different phases based on key features, including subjects, recordings, analysis and alarms, and reference standard.

3 | RESULTS

The database searches yielded a total of 2394 articles, out of which 1537 were screened based on title and abstract and 136 articles were selected for full-text review. After reviewing the full text based on eligibility criteria, 24 articles were included for analysis. Twenty studies evaluated

seizure detection algorithms,^{27–46} whereas the remaining four studies focused on seizure prediction, including forecasting the likelihood of seizures (periods of high and low risk).^{24,47–49} The screening stages and results are outlined in more detail in Figure 2. Articles were excluded mainly due to being a conference abstract or review, the study design not involving the validation of an algorithm on patients, or not using a wearable device to measure cardiovascular signals. Studies that recorded and analyzed cardiovascular signals before or during seizures but not in the context of validating a seizure detection or prediction algorithm were also excluded.

3.1 | Studies on seizure detection

The characteristics of the 20 seizure detection studies included in the review are listed in Table 1. The number of patients included in the analysis for seizure detection ranged between 3 and 94 participants (median = 15.5 participants). Most studies took place in an inpatient setting (n = 17), where patients with a diagnosis of epilepsy admitted to an epilepsy monitoring unit (EMU) for presurgical evaluation, seizure assessment, or diagnostic purposes were recruited. Two studies were conducted in a residential




FIGURE 2 PRISMA (Preferred Reporting Items for Systematic Reviews and Meta-Analyses) flowchart for database search, screening, and selection of studies.

TABLE 1 Study characteristics and wearable devices of studies on seizure detection grouped based on cardiac parameters.

Studies included (Author, year)	Study characteristics			Wearable devices			Physiological signal(s) collected
	Study population	Adults/children	Study setting	Reference standard	Wearable device to collect cardiac data	Location of device	
Heart rate (HR)							
Ali and Alam 2020 ²⁷	N/A	N/A	N/A	N/A	Prototype sensor	Wrist	ACC, PPG, TEMP
Arends et al 2018 ²⁸	Adults with epilepsy ID with a history of >1 major nocturnal seizure per month and resided in a long-term facility of 1 of the participating epilepsy centers	Adults	Inpatient and outpatient	Infrared-sensitive video camera	Nightwatch	Arm	ACC, PPG
Cogan et al 2017 ³¹	Adults admitted to the EMU at Presbyterian Hospital of Dallas, Texas.	Adults	Inpatient	Video-EEG	Nonin WristOx2	Wrist	EDA, PPG, SpO ₂
Masse et al 2013 ⁴¹	Patients previously diagnosed with epileptic seizures, with some forms of HR changes	Adults	Inpatient	Video recording and EMFIT	Custom miniaturized ECG monitor	Chest	ECG
Lazeron et al 2022 ⁴⁰	Children (3-18 years old) with refractory epilepsy and an ID living at home or in a specialized institutional residential care setting	Children	Inpatient and outpatient	Infrared-sensitive video camera	Nightwatch	Arm	ACC, PPG
Bottocher et al 2022 ²⁹	Patients (7-80 years old) with a diagnosis of epilepsy recruited as part of their standard clinical epilepsy care	Adults and children	Inpatient	Video-EEG	Empatica E4	Wrist	ACC, EDA, PPG
Bruno et al 2021 ³⁰	People with focal epilepsy admitted to the EMU at King's College Hospital, London UK for diagnostic reasons or presurgical evaluation	Adults and children	Inpatient	Video-EEG	Bespoke upper arm band	Arm	ECG, ACC, EDA, sEMG
van Andel et al 2017 ⁴⁴	Patients (above 2 years old) with a history of nocturnal seizure frequency >1 seizure/wk, admitted to one of the centers for long term (>24 h) video-EEG monitoring	Adults and children	Inpatient	Video-EEG	Shimmer sensor	Arm	ACC, ECG
Cogan et al 2015 ³²	Patients admitted to the EMU	N/A	Inpatient	EEG	Nonin WristOx2	Wrist	EDA, PPG, SpO ₂
Henze et al 2021 ³⁵	Epilepsy patients under video-EEG monitoring	N/A	Inpatient	Video-EEG	cosinuss° In-Ear sensor	Ear	ACC, PPG
Zsom et al 2019 ⁴⁶	Patients undergoing inpatient video EEG long-term monitoring	N/A	Inpatient	Video-EEG	Empatica E4	Wrist	EDA, PPG
Heart rate variability (HRV)							
Munch Nielsen et al 2022 ⁴²	Patients admitted to the EMU at Zealand University Hospital for diagnostic evaluation	Adults and children	Inpatient	Video-EEG	Bitrium Faros 180°	Chest	ACC, ECG, EEG
Foroughifar et al 2019 ³³	Patients with epilepsy who underwent in-hospital recording of their seizures for diagnostic purposes	N/A	Inpatient	Video-EEG	SmartCardia INYU wearable sensor	Chest	ECG
Blood volume pulse							
Tang et al 2021 ⁴³	Patients admitted to the Boston Children's Hospital EMU	Adults and children	Inpatient	Video-EEG	Empatica E4	Wrist or ankle	ACC, EDA, PPG, TEMP

(Continues)

TABLE 1 (Continued)

Studies included (Author, year)	Study characteristics			Wearable devices			
	Study population	Adults/children	Study setting	Reference standard	Wearable device to collect cardiac data	Location of device	Physiological signal(s) collected
Heart rate and heart rate variability							
Jahanebekam et al 2021 ³⁶	Adult patients aged 18 years or older with refractory epilepsy who underwent video-EEG monitoring	Adults	Inpatient	Video-EEG (non-invasive scalp-EEG recordings)	EcgMove	Chest	ECG, ACC, EDA
Vandecasteele et al 2017 ⁴⁵	Patients with refractory epilepsy who underwent presurgical evaluation at UZ Leuven Gasthuisberg	Adults	Inpatient	Video-EEG	Bitium Faros 180° (ECG) and Empatrica E4 (PPG)	Chest and wrist	ECG
Hegarty-Craver et al 2021 ³⁴	Children (2-17 y) undergoing video-electroencephalogram monitoring for clinical care	Children	Inpatient	Video-EEG	Zephyr Biopatch and Bitium Faros 180°	Chest	ACC, ECG
Jeppesen et al 2017 ³⁸	People (5-76 years old) with a diagnosis of probable focal or generalized epilepsy, enrolled for a long-term video-EEG monitoring in Aarhus University Hospital	Adults and children	Inpatient	Video-EEG	ePatch device	Chest	ECG
Jeppesen et al 2019 ³⁹	Patients who were enrolled for diagnostic reasons or presurgical evaluation in the EMU	Adults and children	Inpatient	Video-EEG	ePatch device	Chest	ECG
Jeppesen et al 2020 ³⁷	Patients who were enrolled for diagnostic reasons or presurgical evaluation in the EMU	Adults and children	Inpatient	Video-EEG	ePatch device	Chest	ECG

Abbreviations: ACC, accelerometry; ECG, electrocardiogram; EDA, electrodermal activity; EEG, electroencephalogram; EMU, epilepsy monitoring unit; HR, heart rate; ID, intellectual disability; iEEG, intracranial electroencephalography; N/A, information not available; PPG, photoplethysmography; sEMG, surface electromyography; SpO₂, blood oxygen saturation; TEMP, temperature.

setting, where they validated seizure detection devices for detecting nocturnal seizures.^{28,40} In studies with inpatient monitoring, patients were allowed to move around freely and perform normal daily activities despite being confined to a hospital room.^{33,42} Fifteen studies used video-electroencephalography (EEG) as a reference to validate seizure events,^{29–31,33–39,42–46} where clinical experts annotated the electrographic seizure onset and offset. The other studies used infrared-sensitive video cameras,^{28,40} EEG without video recording,³¹ and video recording with EMFIT monitor⁴¹ as the reference standard. One study did not report on the reference standard used.²⁷ Some of the studies also reported incomplete data and that not all patients that enrolled in the study were included in the analysis, mainly due to factors such as poor connection or signal quality,^{34,44} withdrawal of participants,²⁸ insufficient or unsuitable seizures,^{29,30,37,39,42,46} unusable data,^{31,32,34} participant's non-compliance to study protocol.⁴⁰

3.1.1 | Wearable devices

Information on wearable devices used in the seizure detection studies is listed in [Table 1](#). A wide range of non-invasive wearable devices were used to collect the ECG or HR data. Four studies used Empatica E4,^{29,43,45,46} three studies used the ePatch device,^{37–39} and three studies used Bittium Faros 180°.^{34,42,45} Other wearable devices that were noted include Nightwatch,^{28,40} Nonin WristOx2,^{31,32} SmartCardia INYU wearable sensor,³³ Zephyr Biopatch,³⁴ cosinuss° In-Ear sensor,³⁵ EcgMove,³⁶ and Shimmer sensor.⁴⁴ Interestingly, two studies developed custom wearable devices to collect the ECG or PPG data.^{30,41} The devices were primarily either worn on the chest (n=8)^{33,34,36–39,41,42} or wrist (n=5).^{27,29,31,32,46} In the remaining studies, the wearable devices were worn on the arm (n=4)^{28,30,40,44} and ear (n=1).³⁵ One study used two wearable devices, where one device was worn on the chest and the other on the wrist,⁴⁵ while another study allowed their participants to wear the device either on their wrists or ankles.⁴³

3.1.2 | Seizure detection algorithms

Information on the seizure detection algorithms and their performance are listed in [Table 2](#). Eleven studies extracted HR features for the seizure detection algorithms, with a combined sensitivity range of 56%-100% and a false alarm rate (FAR) of 0.02-8/h. Most of the studies that used HR as a basis for their seizure detection algorithm (eight out of 11 studies) validated their algorithm retrospectively using an existing dataset.^{27,29–32,35,41,46}

For example, data analysis and algorithm testing were performed after the recording of physiological signals from patients (offline). Of the 11 seizure detection studies based on HR, four studies involved adult participants,^{27,28,31,41} three studies included both adult and child participants^{29,30,44} and one study involved children only.⁴⁰ The remaining three studies did not report information on the participants' ages.^{32,35,46}

Studies with adult participants achieved a combined sensitivity range of 85%-100%. The first study used a unimodal approach, where a custom miniaturized wearable ECG monitor was developed and integrated with a beat-detection algorithm and a real-time epileptic seizure detection algorithm to detect seizures. The device was validated in three patients with epilepsy who had HR changes. Tonic-clonic, generalized tonic, and hypermotor seizures were detected with a mean sensitivity of 75% and PPV of 70%.⁴¹ The remaining three studies with adults used a multimodal algorithm that was validated either prospectively²⁸ or retrospectively.^{27,31} In the prospective study, the multimodal sensor detected a median of 14 seizures, with a median sensitivity of 86%, median positive predictive value (PPV) of 49%, and a false positive (FP) rate of 0.25 per night.²⁸ The authors also demonstrated that HR is a critical modality, as it accounted for 92% of the detection of true positives, whereas the ACC modality accounted for only 8% of detections. One retrospective study developed a seizure detection algorithm by analyzing HR, blood oxygen saturation (SpO₂), and EDA biosignals acquired with a wrist-worn device. The personalized algorithm was able to detect seizures with 100% sensitivity and a FAR of 0.00/h in six out of 10 patients.³¹

In the study focused on HR-based seizure detection among children, an adapted algorithm was developed to reduce false alarms, where the alarm was only triggered when the participant was lying in a horizontal position. This algorithm detected 305 out of 384 seizures (median sensitivity: 93%), with a median PPV of 58% and a false negative alarm rate of 0.02/h.⁴⁰ Studies that included both adults and children did not provide a clear distinction in algorithm performance between the two age groups, although one retrospective study used a leave-one-seizure-out method for evaluation across three patients and reported lower sensitivity and higher FAR in one pediatric patient compared to the other two adult patients (sensitivity: 67% vs 100%, FAR₂₄: 41.52 vs 0.85-17.69).²⁹ In this study, HR features were extracted from the BVP signals and combined with ACC and EDA features. An optimized model was also developed, which could detect focal motor seizures with a mean sensitivity of 75%, a mean FAR of 13.4/24h, and a PPV of 2.1%.²⁹

Two studies have used HRV as a parameter for seizure detection.^{33,42} One retrospective validation study

TABLE 2 Seizure detection algorithms grouped based on cardiac parameters.

Studies included (Author, year)	Type of validation	Total participants recruited	No. of patients analyzed	Adults/children	Modality	Seizure type	Results of algorithm performance
Heart rate (HR)							
Ali and Alam 2020 ²⁷	Retrospective	3	3	Adults	Multimodal	Convulsive seizures	Sensitivity: 85% FAR: 26.09%
Arends et al 2018 ²⁸	Prospective	34	28	Adults	Multimodal	Nocturnal seizures: TCS, GT, HK, OM	Sensitivity: 86% PPV: 49% FNAR: 0.03 /night FPAR: 0.25 /night
Cogan et al 2017 ³¹	Retrospective	20	10	Adults	Multimodal	FOIA, FBTCS, GTCS	3 Sensors (n = 6): Sensitivity: 100%, 100% (P) PPV: 86%, 100% (P) FAR: 0.015/h, 0.000/h (P)
Masse et al 2013 ⁴¹	Retrospective	3	3	Adults	Unimodal	TCS, generalized tonic and hypermotor	Sensitivity: 75% PPV: 70.4%
Lazeron et al 2022 ⁴⁰	Prospective	25	23	Children	Multimodal	Nocturnal seizures: TCS, GT, HK, OM	Sensitivity: 93.2% (mean 85.9%, range 47.4%-100% [95% CI 59.9%-100%] FNAR: 0.02/h (range 0.003-0.12) Median PPV: 58.1% (mean 55.5%, range 1.2%-86.6%)
Bottcher et al 2022 ²⁹	Retrospective	243	6	Adults and children	Multimodal	FBTCS, GTCS	Sensitivity: 75% FAR24: 13.4 PPV: 2.1%
Bruno et al 2021 ³⁰	Retrospective	51	12	Adults and children	Multimodal	Focal motor seizures with impaired awareness and focal motor aware seizures	• Sensitivity (motor seizure impaired awareness): 75% (ECG only), 92% (ECG or sEMG or ACC or EDA) • None of motor aware seizures had ECG or EDA manifestations associated
van Andel et al 2017 ⁴⁴	Retrospective	95	23	Adults and children	Multimodal	Nocturnal seizures: GTC, hypermotor, GT, cluster seizures	HR only: Sensitivity = 56%-71%; FAR = 2.3 /night HR + Mvt: Sensitivity = 71%-87%; FAR = 5.9-6.3/night
Cogan et al 2015 ³²	Retrospective	5	3	N/A	Multimodal	Secondarily generalized and complex partial seizures	Sensitivity: 100%, 100% (P) Specificity: 83%, 100% (P) Accuracy: 92%, 100% (P)
Henze et al 2021 ³⁵	Retrospective	N/A	17	N/A	Multimodal	TCS	Mean detection latency: 13 s FAR: 192/24 h Sensitivity: 0.856-0.91 PPV: 0.016
Zsom et al 2019 ⁴⁶	Retrospective	30	18	N/A	Multimodal	Epileptic seizures, PNES, convulsive and non-convulsive	Accuracy: 78%

TABLE 2 (Continued)

Studies included (Author, year)	Type of validation	Total participants recruited	No. of patients analyzed	Adults/children	Modality	Seizure type	Results of algorithm performance
Heart rate variability (HRV)							
Munch Nielsen et al 2022 ⁴²	Retrospective	30	3	Adults and children	Multimodal	Focal tonic and focal nonmotor seizures	Focal tonic: Sensitivity = 84%, FAR = 8/24 h Focal nonmotor seizures: Sensitivity = 100%, FAR = 13/24 h and 5/24
Forooghifar et al 2019 ³³	Retrospective	N/A	18	N/A	Unimodal	Focal seizures	Sensitivity: 88.7% Specificity: 85.7%
Blood volume pulse							
Tang et al 2021 ⁴³	Retrospective	N/A	94	Adults and children	Multimodal	FBTCSs, GTCSs, focal tonic seizures, focal subclinical seizures, focal automatisms, focal behavior arrest, focal clonic seizures, generalized tonic seizures, and generalized epileptic spasms	<ul style="list-style-type: none"> ACC + BVP data fusion provided the best averaged AUC-ROC performance on average Algorithm 1: AUC-ROC of 0.976 (GTCSs) and 0.932 (FBTCSs) Algorithm 2: ACC + BVP data fusion reached the highest overall AUC-ROC of 0.752 applied to all seizure samples lumped together across all seizure types
Heart rate and heart rate variability							
Jahanbekam et al 2021 ³⁶	Retrospective	N/A	35	Adults	Multimodal	GTCS, FANMS, FOIA, FBTCs	Sensitivity: 67% FP rate: 0.03/h
Vandecasteele et al 2017 ⁴⁵	Retrospective	N/A	11	Adults	Unimodal	FOIA	Sensitivity: 64% (overall: 70%) PPV: 2.03% (overall: 2.15%) FAR: 2.35/h (overall: 2.11/h)
Hegarty-Craver et al 2021 ³⁴	Retrospective	62	18	Children	Multimodal	TCS and focal seizures	Sensitivity: 72% FP rate: 0.04/h
Jeppesen et al 2017 ³⁸	Retrospective	14	14	Adults and children	Unimodal	N/A	Sensitivity: 99.979% PPV: 99.976%
Jeppesen et al 2019 ³⁹	Retrospective	100	43	Adults and children	Unimodal	FBTCS, GTCS, FOIA, FAS	Sensitivity: 93.1% FP rate: 0.04/h
Jeppesen et al 2020 ³⁷	Retrospective	47	11	Adults and children	Unimodal	FBTCS, GTCS, FOIA, FAS	Sensitivity: 87.0% FP rate: 0.38/h

Abbreviations: ACC, accelerometry; AUC-ROC, area under the receiver operating characteristic curve; BVP, blood volume pulse; CI, confidence interval; ECG, electrocardiogram; EDA, electrodermal activity; FAR, false alarm rate; FAS, focal aware seizures; FANMS, focal aware non-motor seizures; FBTCs, focal to bilateral tonic-clonic seizures; FNAR, false negative alarm rate; FOIA, focal onset impaired awareness seizures; FP, false positive; FPAR, false positive alarm rate; GT, generalized tonic; GTCS, generalized tonic-clonic seizures; HK, hyperkinetic seizures; HR, heart rate; HRV, heart rate variability; IoC, improvement over chance; Mvt, movement; N/A, information not available; OM, other major seizures; P, personalized algorithm; PNES, psychogenic non-epileptic seizures; PPV, positive predictive value; sEMG, surface electromyography; TCS, tonic clonic seizures; TIW, time in warning.

collected ECG, ACC, and behind-the-ear EEG signals from both adults and children using separate devices and extracted HRV measures such as Modified Cardiac Sympathetic Index (ModCSI) and Modified Cardiac Sympathetic Index with Slope (ModCSISlope) from the R peaks of the ECG signal.⁴² A support vector machine (SVM) algorithm, another machine learning tool, was used to classify seizure or non-seizure events based on the multimodal signal features extracted. This study reported the detection of focal tonic (sensitivity: 84%, FAR: 8 per 24 h) and focal non-motor seizures (sensitivity: 100%, FAR: 13 per 24 h) in three patients. In another retrospective study, the ECG signal collected using the chest-worn SmartCardia INYU wearable sensor was used to extract the R-R interval (RRI) and ECG-Derived Respiration (EDR) time series. For HRV analysis, time domain, frequency domain, Lorenz plot, and multifractality features were extracted from the RRI to assess changes in cardiac function. The random forest classifier, a machine learning tool, was applied to classify seizure and non-seizure segments. The algorithm was able to detect focal seizures with a sensitivity of 88.7% and a specificity of 85.7%.³³

Six studies evaluated seizure detection based on both HR and HRV parameters.^{34,36–39,45} Three of these studies included both adult and child participants, two involved adults only, while the remaining study included children only. Among the studies with adult participants, one unimodal algorithm study analyzed and compared ECG and PPG wearable devices for seizure detection in patients with temporal lobe epilepsy (TLE).⁴⁵ The seizure detection algorithm in this study utilized HRV and pulse rate variability extracted from ECG and PPG signals, respectively, and identified an HR increase before performing classification using a SVM classifier. The wearable ECG achieved the highest sensitivity (70%) compared to the hospital ECG and wearable PPG, with a comparable FAR of 2.11/h. The other study with adult participants retrospectively validated a multimodal algorithm combining HRV and HR with ACC and EDA, achieving a sensitivity of 67% and an FP rate of 0.03/h.³⁶

In the study with children, a multimodal seizure detection algorithm based on ECG and ACC signals was evaluated.³⁴ The authors also developed a custom prototype unit to employ the algorithm and allow for real-time seizure detection. The cardiac algorithm, where both HR and HRV were analyzed, was able to detect tonic-clonic and focal seizures without focal to bilateral tonic-clonic features with an overall sensitivity of 72% and FAR of 0.04/h. Interestingly, when ECG and ACC parameters were combined, four seizures were detected faster, but the overall sensitivity did not improve.³⁴ Studies that evaluated HRV and HR-based algorithms in both adults and children

($n = 3$) were all retrospective studies evaluating unimodal algorithms that achieved a combined sensitivity range of 87%–100%. However, differences in performance between adults and children were not reported.^{37–39}

3.2 | Studies on seizure prediction

Characteristics and wearable devices of studies on seizure prediction are listed in Table 3, and information on their algorithms is summarized in Table 4. Three studies employed multimodal seizure prediction algorithms and conducted retrospective validation.^{24,47,48} The first retrospective study evaluated a multimodal algorithm combining EDA, HR, and skin temperature collected with the wrist-worn Empatica E4 in three patients with refractory epilepsy. A naïve Bayes classifier was trained on a set of sample data and then evaluated using five-fold cross-validation for preictal and interictal classification during wakefulness, achieving a sensitivity of 78% and a specificity of 80%.⁴⁷ The second retrospective study also used Empatica E4 to acquire ACC, EDA, PPG, and temperature data in adults and children with epilepsy. Out of 69 patients included in the analysis, seizure forecasting was significantly better than chance for 43.5%, of which achieved a mean improvement of chance (IoC) of $28.5 \pm 2.6\%$, a mean sensitivity of $75.6 \pm 3.8\%$, and a mean percentage of time spent in warning (TiW) of $47.2 \pm 3.4\%$ (mean \pm SEM). The same study also assessed the effect of reducing the training dataset on seizure forecasting algorithm performance and reported improvements in IoC with larger datasets.²⁴ Another retrospective study utilized a smartwatch to acquire PPG, sleep, and step count data, and a smartphone seizure diary app to monitor seizures in adults with refractory epilepsy. HR, HR cycles, and HRV features were extracted from the PPG signal for the algorithm to forecast periods of high and low risk of seizures. The hourly forecasts achieved a median accuracy of 86%, and the average time spent in high-risk (prediction time) prior to a seizure onset was 37 minutes. Meanwhile, the daily forecast achieved 83% median accuracy, and the average prediction time before a seizure was 3 days.⁴⁸

Only one study prospectively validated their seizure prediction algorithm and extracted both time-domain and frequency-domain HRV features.⁴⁹ A custom wearable ECG device for seizure prediction was developed in this study, which consists of an RR interval telemeter connected to a custom smartphone app via Bluetooth connection. The smartphone app is able to receive and analyze RRI, which is used to extract HRV features. A machine learning tool known as multivariate statistical process control for seizure prediction was employed, where a successful prediction was defined as a seizure identified

TABLE 3 Study characteristics and wearable devices of studies on seizure prediction.

Study characteristics			Wearable devices				
Studies included (Author, year)	Study population	Adults/children	Study setting	Reference standard	Wearable device to collect cardiac data	Location of device	Physiological signal(s) collected
Stirling et al 2021 ⁴⁸	Adults (18y and over) with a confirmed epilepsy diagnosis	Adults	Outpatient	Seizure events manually reported in a Smartphone diary app	FitBit	Wrist	PPG, sleep, step counts
Meisel et al 2020 ²⁴	Patients (2-22 years old) with epilepsy admitted to the long-term video-EEG monitoring	Adults and children	Inpatient	Video-EEG	Empatica E4	Wrist	ACC, EDA, PPG, TEMP
Yamakawa et al 2020 ⁴⁹	Patients with refractory epilepsy admitted and underwent clinical video-EEG monitoring for presurgical evaluation or seizure assessment	Adults and children	Inpatient	Video-EEG	Custom wearable ECG device	Chest	ECG
Al-Bakri et al 2018 ⁴⁷	Patients admitted for invasive presurgical evaluation at the University of Kentucky Medical Center	N/A	Inpatient	Video-EEG and iEEG	Empatica E4	Wrist	EDA, PPG, TEMP

Abbreviations: ACC, accelerometry, ECG, electrocardiogram, EDA, electrodermal activity, EEG, electroencephalogram, iEEG, intracranial electroencephalogram, N/A, information not available, PPG, photoplethysmography, TEMP, temperature.

TABLE 4 Seizure prediction algorithms.

Studies included (Author, year)	Type of validation	Total participants recruited	No. of patients analyzed	Adults/children	Modality	Cardiac parameter used for algorithm	Seizure type	Results of algorithm performance	
								Median accuracy (hourly forecast): 86%	Median accuracy (daily forecast): 83%
Stirling et al 2021 ⁴⁸	Retrospective	39	11	Adults	Multimodal	HR cycles, HRV, HR	N/A	Median accuracy (hourly forecast): 86%	Median accuracy (daily forecast): 83%
Meisel et al 2020 ²⁴	Retrospective	317	69	Adults and children	Multimodal	BVP	Primary and secondary generalized, and focal seizures	Mean IoC = 28.5 ± 2.6%	Mean sensitivity = 75.6 ± 3.8%
Yamakawa et al 2020 ⁴⁹	Prospective	14	7	Adults and children	Unimodal	HRV	FIAS, FBTCs, FAS	Mean TIW (ie, the percentage of time spent in warning) = 47.2 ± 3.4% (mean ± SEM)	Sensitivity: 85.7%
Al-Bakri et al 2018 ⁴⁷	Retrospective	3	3	N/A	Multimodal	HR	N/A	FP rate: 0.62/h	Sensitivity = 78%
									Specificity = 80%
									Cohen's kappa = 55%

Abbreviations: BVP, blood volume pulse, FAS, focal aware seizures, FBTCs, focal to bilateral tonic-clonic seizures, FIAS, focal impaired awareness seizures, FP, false positive, FPAR, false positive alarm rate, HR, heart rate, HRV, heart rate variability, IoC, improvement over chance, N/A, information not available, SEM, standard error of mean, TIW, time in warning.

between 15 minutes and immediately before seizure onset. The custom seizure prediction system demonstrated the ability to predict focal impaired awareness seizures, focal to bilateral tonic-clonic seizures, and focal aware seizures in both adults and children (sensitivity: 85.75%, FAR: 0.62/h).⁴⁹

3.3 | Quality assessment

The studies included in this review are of mixed quality (Table 5). Some studies did not provide a clear description of the recruitment process or reference standard; therefore, it was difficult to assess the limitations and quality of these studies. Although a majority of the studies used EEG or video-EEG as the reference standard ($n = 19$), only five studies^{28,37,39,44,45} explicitly reported blinded annotation of seizures. It was unclear in the remaining studies whether the reference standard was reviewed without knowledge of the cardiac data. In the flow and timing domain, studies mostly had a low risk of bias as the data from the reference standard was collected concurrently with the cardiac data, and all patients received the same reference standard. Due to the heterogeneity in study design and algorithm performance indicators reported in the studies, we could not conduct a meta-analysis in this present study.

Most of the studies are categorized as phase 1 ($n = 19$) according to the proposed standards by Beniczky and Ryvlin (Table 6). Three studies are categorized as phase 2,^{28,37,39} while the remaining two studies are phase 0.^{30,48} Although all studies used a dedicated device and a majority used video recording or video-EEG ($n = 22$) as the reference standard, some studies could not be classified as phase 2 due to an inadequate number of patients or because the safety of the device was not addressed. Thirteen studies trained and tested their algorithms on the dataset, and 10 studies used predefined algorithms and cutoff values. Only four studies evaluated their algorithms in real time.^{28,40,41,49}

4 | DISCUSSION

The findings from this systematic review highlighted both the promise and challenges of the feasibility of cardiac-based seizure detection and prediction using non-invasive wearable devices. There is a clear feasibility with utilizing cardiac-based algorithms in non-invasive wearable devices for seizure detection, especially in adult populations of epilepsy with generally good cardiovascular health; however, the feasibility of them for seizure prediction, whether in adults or children, may be too soon to conclude given the lack of data/clinical studies on their real-world prospective usage. Moreover, the reliability, validity, and

sensitivity of these devices when taking into account the effect of cardiovascular abnormalities, such as a history of bradycardia or tachycardia, among people with epilepsy have not been investigated. The findings were based on articles that were either in phase 1 (proof-of-principle) or phase 2 (safety of device addressed) of their study, with none being in phase 3 (confirmation of safety and accuracy) or phase 4 (in-field, usability aspects), suggesting that the feasibility of these cardiac-based seizure detection and prediction devices is still in the early stages of development and validation, mainly in a controlled environment. Real-world effectiveness accounting for patient clinical heterogeneity in not only seizure development but also general health, as well as patient usability in their daily lives, still requires further research. Since the risk of bias was found to be mostly unclear for a number of the studies due to a lack of information in the reference standards used, the conclusions from these studies could be potentially biased. Although a dedicated device was used in all studies, most of them lack an assessment of device safety and even have low sample sizes, contributing to a lack of feasibility information and data interpretation biasness. Nevertheless, findings from these studies may still contribute as an important stepping stone toward the epilepsy diagnostic and possibly therapeutic avenue upon further validation.

Some of the studies have trained and validated their own machine learning algorithms to detect changes in physiological signals that indicate seizure events,⁴³ while others have also used pre-trained algorithms. It is unclear at this point in time whether there are any confounding variables to be considered with either one style of algorithm but a recent paper stated that while pre-trained models may speed up optimization of the algorithm, they may have biased notions from previous machine learning training that will lead to inaccuracies in current data interpretation.⁵⁰ Moreover, multiple different machine learning techniques have been employed for seizure detection and prediction across the studies, resulting in diverse performance results, thereby creating a more convoluted conclusion. However, the diversity in machine learning techniques used may bring to light the limitations and advantages of each technique, thereby providing future studies with a better idea of which technique would be best suited for further testing.

In addition, diverse algorithmic performance was also reported when comparing seizure detection between adults and children. Among studies that use HR as an input for the seizure detection algorithm, a generally higher FAR was observed in children. However, one prospective study involving children only achieved an improvement in FAR by using an adapted algorithm.⁴⁰ In studies that used HRV and HR, higher detection

TABLE 5 Quality assessment of the studies included using QUADAS-2 tool.

Studies included (Author, year)	Risk of bias				Applicability concerns			
	Patient selection	Index test	Reference standard	Flow and timing	Patient selection	Index test	Reference standard	Reference standard
Seizure detection								
Ali and Alam 2020 ²⁷	◆	◆	◆	◆	◆	◆	◆	◆
Arends et al 2018 ²⁸	◆	◆	◆	◆	◆	◆	◆	◆
Botlicher et al 2022 ²⁹	◆	◆	◆	◆	◆	◆	◆	◆
Bruno et al 2021 ³⁰	◆	◆	◆	◆	◆	◆	◆	◆
Cogan et al 2015 ³²	◆	◆	◆	◆	◆	◆	◆	◆
Cogan et al 2017 ³¹	◆	◆	◆	◆	◆	◆	◆	◆
Forooghifar et al 2019 ³³	◆	◆	◆	◆	◆	◆	◆	◆
Hegarty-Craver et al 2021 ³⁴	◆	◆	◆	◆	◆	◆	◆	◆
Henze et al 2021 ³⁵	◆	◆	◆	◆	◆	◆	◆	◆
Jahanbekam et al 2021 ³⁶	◆	◆	◆	◆	◆	◆	◆	◆
Jeppesen et al 2017 ³⁸	◆	◆	◆	◆	◆	◆	◆	◆
Jeppesen et al 2019 ³⁹	◆	◆	◆	◆	◆	◆	◆	◆
Jeppesen et al 2020 ³⁷	◆	◆	◆	◆	◆	◆	◆	◆
Lazeron et al 2022 ⁴⁰	◆	◆	◆	◆	◆	◆	◆	◆
Masse et al 2013 ⁴¹	◆	◆	◆	◆	◆	◆	◆	◆
Munch Nielsen et al 2022 ⁴²	◆	◆	◆	◆	◆	◆	◆	◆
Tang et al 2021 ⁴³	◆	◆	◆	◆	◆	◆	◆	◆
van Andel et al 2017 ⁴⁴	◆	◆	◆	◆	◆	◆	◆	◆
Vandecasteele et al 2017 ⁴⁵	◆	◆	◆	◆	◆	◆	◆	◆
Zsom et al 2019 ⁴⁶	◆	◆	◆	◆	◆	◆	◆	◆
Seizure prediction								
Al-Bakri et al 2018 ⁴⁷	◆	◆	◆	◆	◆	◆	◆	◆
Meisel et al 2020 ²⁴	◆	◆	◆	◆	◆	◆	◆	◆
Stirling et al 2021 ⁴⁸	◆	◆	◆	◆	◆	◆	◆	◆
Yamakawa et al 2020 ⁴⁹	◆	◆	◆	◆	◆	◆	◆	◆

◆ = High risk of bias; ◆ = Unclear risk of bias; ◆ = Low risk of bias.

TABLE 6 Assessment of included studies according to proposed standards by Beniczky and Ryvlin.

Studies included (Author, year)	Study phase ^a	Subjects		Recordings			Analysis and alarms			Reference standard				
		Simulation/ healthy subjects	No. of pts with seizures	No. of seizures	Conventional methods	Dedicated device	Continuous	MultiCentre	Offline/ Retrospective	Training & testing using the dataset	Predefined algorithm and cutoff values	Real time	Blinded	Video or video-EEG recordings
Seizure detection														
Ali and Alam 2020 ²⁷	1	+	1-10	15-30	-	+	-	+	-	-	+	-	-	-
Arends et al 2018 ²⁸	2	-	20-50	≥75	-	+	+	-	-	+	+	+	+	-
Bottcher et al 2022 ²⁹	1	-	1-10	15-30	-	+	+	+	+	-	-	-	+	-
Bruno et al 2021 ³⁰	0	-	10-20	30-75	+	+	+	-	-	-	-	-	+	-
Cogan et al 2015 ³²	1	-	1-10	1-15	-	+	-	+	+	-	-	-	+	-
Cogan et al 2017 ³¹	1	-	1-10	1-15	-	+	-	+	+	-	-	-	+	-
Forooghifar et al 2019 ³³	1	-	10-20	≥75	+	+	-	+	+	-	-	-	+	-
Hegarty-Craver et al 2021 ³⁴	1	-	10-20	15-30	-	+	-	+	+	-	-	-	+	-
Henze et al 2021 ³⁵	1	-	10-20	15-30	-	+	-	+	+	-	-	-	+	-
Jahnbekam et al 2021 ³⁶	1	-	20-50	30-75	-	+	-	+	+	-	-	-	+	-
Jeppesen et al 2017 ³⁸	1	-	10-20		-	+	-	+	+	-	-	-	+	-
Jeppesen et al 2019 ³⁹	2	-	20-50	≥75	-	+	+	+	-	+	+	+	+	-
Jeppesen et al 2020 ³⁷	2	-	10-20	30-75	-	+	+	+	-	+	+	+	+	-
Lazeron et al 2022 ⁴⁰	1	-	20-50	≥75	-	+	-	+	-	+	+	+	+	+
Masse et al 2013 ⁴¹	1	+	1-10	15-30	-	+	-	+	+	-	+	+	+	-
Munch Nielsen et al 2022 ⁴²	1	-	1-10	30-75	-	+	-	+	+	-	+	-	+	-
Tang et al 2021 ⁴³	1	-	≥50	≥75	-	+	-	+	+	-	+	-	+	-
van Andel et al 2017 ⁴⁴	1	-	20-50	≥75	-	+	+	+	-	+	+	+	+	-
Vandecasteele et al 2017 ⁴⁵	1	-	10-20	30-75	+	+	-	+	+	-	+	+	+	-
Zsom et al 2019 ⁴⁶	1	-	10-20	30-75	-	+	-	+	+	-	-	-	+	-
Seizure prediction														
Al-Bakri et al 2018 ⁴⁷	1	-	1-10	N/A	+	+	-	+	+	-	+	-	+	-
Meisel et al 2020 ⁴⁴	1	-	≥50	≥75	-	+	-	+	+	-	+	-	+	-
Stirling et al 2021 ⁴⁸	0	-	10-20	≥75	-	+	-	+	+	-	+	-	-	+
Yamakawa et al 2020 ⁴⁹	1	+	1-10	1-15	+	+	+	+	-	+	+	+	+	-

Abbreviations: N/A, information not available; No., number; pt, patient.

^aStudies are categorized into the following phases: Phase 0: initial studies for starting or developing a novel method. Phase 1: proof-of-principle studies. Phase 2: studies using dedicated seizure detection device with safety of the device addressed. Phase 3: studies on final confirmation of safety and accuracy. Phase 4: in-field studies of seizure detection devices patients home environments, addressing usability aspects.

sensitivity was reported among children compared to the studies involving adults, possibly due to the utilization of a patient-dependent algorithm in the study with children. It is difficult to compare the performance between adults and children among the studies that reported on both, as they did not provide a clear difference between the two age groups. Resting HR was found to increase with age⁵¹ and a pre-ictal decrease in HR has been reported exclusively in studies on pediatric population,¹⁸ therefore, the large variation in HR among different age groups may contribute to diverse performance results. While training and testing algorithms on one specific age group at a time could potentially improve the detection or prediction performance of the cardiac-based devices, it should also be noted that adults and children have distinct seizure profiles and requirements and therefore should always be treated as separate subject groups in future studies, unlike some studies reviewed in this manuscript.^{24,29,30,43,44,49}

In regards to the cardiac parameters utilized, it was found that HR was most commonly extracted and analyzed for seizure detection, with a few studies reporting the use of HRV or the combination of HRV and HR. There is a high incidence of pre-ictal HR increase more specifically in studies involving TLE patients, adults, or patients receiving antiseizure medications (ASM).¹⁸ However, there may be limitations, as not all seizures have changes in HR⁵² and may be prone to fluctuations contributed by medication, stress, age, sleep quality, and exercise.⁵³ Some studies have provided a possible solution to this by using a multimodal algorithm and comparing it with a unimodal algorithm,^{24,30,44} and others have also asked patients to perform an exercise or stress test to sample real-life situations.^{37,39} Another meaningful cardiac measurement is the HRV, and studies that used this parameter as a basis for seizure detection achieved a slightly higher sensitivity compared to those using HR only. HRV is regulated by the balance between the sympathetic and parasympathetic nervous systems; therefore, changes in HRV serve as an indicator of the ANS function. Studies investigating the correlation between interictal HRV and epileptic seizures reported a lower HRV, suggesting an imbalance that shifts more toward sympathetic activity.⁵⁴ This is in line with another study, which also reported an increase in peri-ictal (pre-ictal and post-ictal) sympathetic activity in generalized tonic-clonic seizures.⁵⁵ Taken together, it is clear that both HR and HRV cardiac parameters should be included in the device algorithm to improve the sensitivity and specificity of seizure detection and prediction, which could also greatly improve with the addition of other physiological parameters such as skin temperature and electrodermal activity. The combination of cardiac

and other physiological parameters may also reduce the chances of false alarms, which will encourage greater patient compliance and usability.

Unfortunately, detection of seizures was more frequently assessed and reported among the studies compared to seizure prediction. Since the latter is more aimed at notifying patients and caregivers of an imminent seizure,⁵⁶ it is imperative to accurately determine pre-ictal periods when training the algorithm.²⁴ Accurate prediction of seizures will allow patients and caregivers to take the necessary precautions (medications or safe space) prior to their seizures, thus eliminating possible scenarios that may reduce their quality of life. Among the seizure prediction studies, there were variations in the performance metrics reported, maybe due to the different forms of prediction evaluated. Some studies provided a binary prediction (yes or no), while others forecast periods of high or low seizure likelihood, which could present more benefits as it allows users to plan activities and manage treatment according to the different periods of seizure likelihood.^{11,12} At present, it is difficult to compare prediction performance between age groups as we did not find any studies that evaluated seizure prediction exclusively in pediatric patients. The studies included either adults only or both adults and children. In the studies that included both adults and children, they did not provide separate performance metrics for either population. Similar to seizure detection studies, multimodal algorithms or patient-specific algorithms may help to improve prediction rates, especially since there is no one-size-fits-all approach to developing a seizure prediction algorithm.⁵⁷ Billeci et al⁵⁸ developed a patient-specific algorithm for seizure prediction and found that optimal performance was achieved in patients with more conventional seizures. A considerable number of studies have also used multiple modalities, combining cardiac parameters with other physiological data in an effort to improve algorithm performance. Nevertheless, one or two modalities could be sufficient, depending on the type of seizures or the presence of ictal tachycardia.⁴²

Wearable technology has made a remarkable impact in healthcare by allowing non-invasive monitoring of patients' health status and providing easier access to information for physicians. The studies included in the analysis have used a wide range of wearable devices, collecting multiple physiological signals that are utilized for seizure detection or prediction algorithms. Most studies used devices that are currently available in the market, and studies that have developed custom wearable devices are currently at the prototype stage and have reported preliminary data. Further clinical testing, particularly on validity and reliability, as well as an evaluation of user acceptance, may still be needed. For instance, a study investigating signal quality in wearable

devices used for epilepsy management and monitoring has evaluated the patient experience and revealed their significant preference for using wrist-worn devices.⁵⁹ Despite the ease and convenience associated with wearable devices, motion artifacts caused by normal daily activities should also be taken into consideration. Signal quality may also differ between individuals due to device or battery failures, consequently resulting in a lack of usable data for analysis. Yamakawa and colleagues have suggested that an ECG that can be worn like clothing may be an option to improve signal and reduce motion artifacts.⁴⁹ Nevertheless, most of the currently available studies were conducted in an inpatient setting, where data from wearable sensors was collected either prospectively or from an existing dataset, and algorithms were validated retrospectively. Hence, the real-world factors such as motion artifacts and battery failures (loss of signal) that could influence the sensitivity of the seizure prediction still lack clarity.

This systematic review is limited by the lack of statistical analysis or meta-analysis to objectively compare the different cardiac parameters used in the seizure detection and prediction algorithms. This is due to the large heterogeneity in study design, setting, and population among the included studies, representing a challenge that may be overcome in the future by following guidelines for conducting and reporting seizure detection or prediction studies.^{26,60,61} Developing studies using these guidelines will ensure that studies are comparable and data can be shared across different seizure detection or prediction research groups, subsequently improving the quality of evidence.

Based on current advancements in technology and digital health, there is a possibility that patient-specific algorithms with an integration of multiple physiological parameters that enhance the accuracy and reliability of cardiac-based seizure detection and prediction devices may be available in the near future. Indeed, real-time validation is first required, especially for the seizure prediction device, to ensure patient compliance and acceptance do not confound the validity and reliability of the seizure prediction. Moreover, with real-time clinical studies, preferably long-term studies, the safety, cost-effectiveness, logistics, and practical utility of the devices can be assessed as well. Large-scale and long-term patient data are required to develop and refine patient-specific algorithms. In addition, prospective validation of the algorithms in a real-world setting and assessment of signal quality would also be useful, taking into account any artifacts and noise that could be contributed by normal daily activities. A recent systematic review discovered that performance, design, comfort, and cost are crucial factors that determine the acceptance of wearable devices in real-world settings although this was not specific to seizure detection or prediction.⁶² Additionally, people

with epilepsy highly prefer non-stigmatizing devices that can be seamlessly integrated into their daily lives thereby justifying the need for real-world usability studies for these cardiac-based seizure detection and prediction devices.⁶¹

Once validated, the cardiac-based seizure device, particularly the seizure prediction device, will be a game-changer in epilepsy management, as treatment against seizures can be utilized more efficiently in a proactive manner than the current reactive seizure management strategies, thereby ensuring timely prevention of seizures and reducing the occurrence of drug adverse effects and resistance caused by overloading of current ASMs. In fact, by utilizing the cardiac-based seizure detection and prediction device, treatment against seizures could also become more automated, leaving children with epilepsy to be more independent in managing their condition and adults to have better adherence to their treatment plan. Thus, successful implementation of these cardiac-based tools into clinical practice may improve current methods of epilepsy management, possibly preventing seizures before their manifestation, thereby ensuring the preservation of quality of life among people with epilepsy.

5 | CONCLUSION

Altogether, the studies analyzed in this systematic review have collectively demonstrated the feasibility of utilizing cardiac parameters as a tool for seizure detection or prediction. The integration of machine learning tools and non-invasive wearable devices signifies a promising advancement in epilepsy care and management. However, future research should focus on refining the detection or prediction performance and providing stronger evidence with more large-scale, multicenter studies conducted in an outpatient, real-life setting. Evaluation of user experience and feedback would be equally important to provide more insight into the clinical value of seizure detection or prediction using non-invasive wearable devices.

AUTHOR CONTRIBUTIONS

All authors have contributed to the preparation of this manuscript. EAS performed the literature search, critical analysis of the articles, and drafted the manuscript; HHMY and MFS performed the literature screening and selection; IWN performed the quality analysis of the literature; JW, JX, AA, CSK, AK, and MFS conceptualized, reviewed, edited, and approved the final manuscript.

ACKNOWLEDGMENTS

Open access publishing facilitated by Monash University, as part of the Wiley - Monash University agreement via the Council of Australian University Librarians.

FUNDING INFORMATION

This project is funded by Monash University Malaysia NEED (NEtwork for Equity through Digital Health) Grant Scheme 2020 (MED/NEED/11-2020/002). EAS, IWN and HHMY are supported by Monash University Malaysia Graduate Research Excellence Scholarship.

CONFLICT OF INTEREST STATEMENT

None of the authors have any conflict of interest to disclose.

DATA AVAILABILITY STATEMENT

Data sharing is not applicable to this article as no new data were created or analyzed in this study.

ETHICS STATEMENT

We confirm that we have read the Journal's position on issues involved in ethical publication and affirm that this report is consistent with those guidelines.


ORCID

Eryse Amira Seth  <https://orcid.org/0000-0002-7827-0807>

Jessica Watterson  <https://orcid.org/0000-0003-0619-0661>

Jue Xie  <https://orcid.org/0000-0003-4549-1627>

Alina Arulsamy  <https://orcid.org/0000-0003-0993-5622>

Hadri Hadi Md Yusof  <https://orcid.org/0000-0002-6702-1952>

Irma Wati Ngadimon  <https://orcid.org/0000-0002-0914-8058>

Ching Soong Khoo  <https://orcid.org/0000-0002-6756-0411>

Amudha Kadirvelu  <https://orcid.org/0000-0001-5646-294X>

Mohd Farooq Shaikh  <https://orcid.org/0000-0001-9865-6224>

REFERENCES

- Nguyen R, Tellez Zenteno JF. Injuries in epilepsy: a review of its prevalence, risk factors, type of injuries and prevention. *Neurol Int.* 2009;1(1):e20.
- Stafstrom CE, Carmant L. Seizures and epilepsy: an overview for neuroscientists. *Cold Spring Harb Perspect Med.* 2015;5(6):1-19.
- Fiest KM, Sauro KM, Wiebe S, Patten SB, Kwon CS, Dykeman J, et al. Prevalence and incidence of epilepsy: a systematic review and meta-analysis of international studies. *Neurology.* 2017;88(3):296-303.
- Ryvlin P, Cross JH, Rheims S. Epilepsy surgery in children and adults. *Lancet Neurol.* 2014;13(11):1114-26.
- Brinkmann BH, Karoly PJ, Nurse ES, Dumanis SB, Nasser M, Viana PF, et al. Seizure diaries and forecasting with wearables: epilepsy monitoring outside the clinic. *Front Neurol.* 2021;12:690404.
- Erba G, Bianchi E, Giussani G, Langfitt J, Juersivich A, Beghi E. Patients' and caregivers' contributions for differentiating epileptic from psychogenic nonepileptic seizures. Value and limitations of self-reporting questionnaires: a pilot study. *Seizure.* 2017;53:66-71.
- Hoppe C, Poepel A, Elger CE. Epilepsy: accuracy of patient seizure counts. *Arch Neurol.* 2007;64(11):1595-9.
- Nickels KC, Zaccariello MJ, Hamiwka LD, Wirrell EC. Cognitive and neurodevelopmental comorbidities in paediatric epilepsy. *Nat Rev Neurol.* 2016;12(8):465-76.
- Ramgopal S, Thome-Souza S, Jackson M, Kadish NE, Sánchez Fernández I, Klehm J, et al. Seizure detection, seizure prediction, and closed-loop warning systems in epilepsy. *Epilepsy Behav.* 2014;37:291-307.
- Dumanis SB, French JA, Bernard C, Worrell GA, Fureman BE. Seizure forecasting from idea to reality. Outcomes of the my seizure gauge epilepsy innovation institute workshop. *eNeuro.* 2017;4(6):ENEURO.0349-17.2017.
- Baud MO, Rao VR. Gauging seizure risk. *Neurology.* 2018;91(21):967-73.
- Stirling RE, Cook MJ, Grayden DB, Karoly PJ. Seizure forecasting and cyclic control of seizures. *Epilepsia.* 2021;62(Suppl 1):S2-S14.
- Beniczky S, Polster T, Kjaer TW, Hjalgrim H. Detection of generalized tonic-clonic seizures by a wireless wrist accelerometer: a prospective, multicenter study. *Epilepsia.* 2013;54(4):e58-61.
- Milosevic M, van de Vel A, Bonroy B, Ceulemans B, Lagae L, Vanrumste B, et al. Automated detection of tonic-clonic seizures using 3-D accelerometry and surface electromyography in pediatric patients. *IEEE J Biomed Health Inform.* 2016;20(5):1333-41.
- Patterson AL, Mudigoudar B, Fulton S, McGregor A, Poppel KV, Wheless MC, et al. SmartWatch by SmartMonitor: assessment of seizure detection efficacy for various seizure types in children, a large prospective single-center study. *Pediatr Neurol.* 2015;53(4):309-11.
- Bottcher S, Bruno E, Manyakov NV, Epatashvili N, Claes K, Glasstetter M, et al. Detecting tonic-clonic seizures in multimodal biosignal data from wearables: methodology design and validation. *JMIR Mhealth Uhealth.* 2021;9(11):e27674.
- Thijs RD, Ryvlin P, Surges R. Autonomic manifestations of epilepsy: emerging pathways to sudden death? *Nat Rev Neurol.* 2021;17(12):774-88.
- Bruno E, Biondi A, Richardson MP, Consortium R-C. Pre-ictal heart rate changes: a systematic review and meta-analysis. *Seizure.* 2018;55:48-56.
- Eggleston KS, Olin BD, Fisher RS. Ictal tachycardia: the head-heart connection. *Seizure.* 2014;23(7):496-505.
- Sivathamboo S, Constantino TN, Chen Z, Sparks PB, Goldin J, Velakoulis D, et al. Cardiorespiratory and autonomic function in epileptic seizures: a video-EEG monitoring study. *Epilepsy Behav.* 2020;111:107271.
- Pavei J, Heinzen RG, Novakova B, Walz R, Serra AJ, Reuber M, et al. Early seizure detection based on cardiac autonomic regulation dynamics. *Front Physiol.* 2017;8:765.
- Bruno E, Simblett S, Lang A, Biondi A, Odoi C, Schulze-Bonhage A, et al. Wearable technology in epilepsy: the views of patients, caregivers, and healthcare professionals. *Epilepsy Behav.* 2018;85:141-9.

23. Vegesna A, Tran M, Angelaccio M, Arcona S. Remote patient monitoring via non-invasive digital technologies: a systematic review. *Telemed J E Health*. 2017;23(1):3–17.
24. Meisel C, El Atrache R, Jackson M, Schubach S, Ufongene C, Loddenkemper T. Machine learning from wristband sensor data for wearable, noninvasive seizure forecasting. *Epilepsia*. 2020;61(12):2653–66.
25. Whiting PF, Rutjes AW, Westwood ME, Mallett S, Deeks JJ, Reitsma JB, et al. QUADAS-2: a revised tool for the quality assessment of diagnostic accuracy studies. *Ann Intern Med*. 2011;155(8):529–36.
26. Beniczky S, Ryvlin P. Standards for testing and clinical validation of seizure detection devices. *Epilepsia*. 2018;59(Suppl 1):9–13.
27. Ali SN, Alam MJ. editorsDevelopment of a wearable 3-risk factor accumulated epileptic seizure detection system with IoT based warning alarm. 2020 IEEE 5th International Conference on Computing Communication and Automation (ICCCA) 2020 30–31 Oct. 2020.
28. Arends J, Thijs RD, Gutter T, Ungureanu C, Cluitmans P, van Dijk J, et al. Multimodal nocturnal seizure detection in a residential care setting: a long-term prospective trial. *Neurology*. 2018;91(21):e2010–9.
29. Bottcher S, Bruno E, Epitashvili N, Dumpelmann M, Zabler N, Glasstetter M, et al. Intra- and inter-subject perspectives on the detection of focal onset motor seizures in epilepsy patients. *Sensors*. 2022;22(9):1–20.
30. Bruno E, Biondi A, Richardson MP, Consortium O. Digital semiology and time-evolution pattern of bio-signals in focal onset motor seizures. *Seizure*. 2021;87:114–20.
31. Cogan D, Birjandtalab J, Nourani M, Harvey J, Nagaraddi V. Multi-biosignal analysis for epileptic seizure monitoring. *Int J Neural Syst*. 2017;27(1):1650031.
32. Cogan D, Nourani M, Harvey J, Nagaraddi V. Epileptic seizure detection using wristworn biosensors. *Annu Int Conf IEEE Eng Med Biol Soc*. 2015;2015:5086–9.
33. Forooghifar F, Aminifar A, Cammoun L, Wisniewski I, Ciumas C, Ryvlin P, et al. A self-aware epilepsy monitoring system for real-time epileptic seizure detection. *Mobile Networks Appl*. 2019;27(2):677–90.
34. Hegarty-Craver M, Kroner BL, Bumbut A, DeFilipp SJ, Gaillard WD, Gilchrist KH. Cardiac-based detection of seizures in children with epilepsy. *Epilepsy Behav*. 2021;122:108129.
35. Henze J, Houta S, Surges R, Kreuzer J, Bisgin P. editorsMultimodal detection of tonic-clonic seizures based on 3D acceleration and heart rate data from an in-ear sensor. *Pattern recognition ICPR international workshops and challenges*. Cham: Springer Science and Business Media Deutschland GmbH; 2021.
36. Jahanbekam A, Baumann J, Nass RD, Bauchhage C, Hill H, Elger CE, et al. Performance of ECG-based seizure detection algorithms strongly depends on training and test conditions. *Epilepsia Open*. 2021;6(3):597–606.
37. Jeppesen J, Fuglsang-Frederiksen A, Johansen P, Christensen J, Wustenhagen S, Tankisi H, et al. Seizure detection using heart rate variability: a prospective validation study. *Epilepsia*. 2020;61(Suppl 1):S41–6.
38. Jeppesen J, Beniczky S, Fuglsang Frederiksen A, Sidenius P, Johansen P. Modified automatic R-peak detection algorithm for patients with epilepsy using a portable electrocardiogram recorder. *Annu Int Conf IEEE Eng Med Biol Soc*. 2017;2017:4082–5.
39. Jeppesen J, Fuglsang-Frederiksen A, Johansen P, Christensen J, Wustenhagen S, Tankisi H, et al. Seizure detection based on heart rate variability using a wearable electrocardiography device. *Epilepsia*. 2019;60(10):2105–13.
40. Lazeron RHC, Thijs RD, Arends J, Gutter T, Cluitmans P, van Dijk J, et al. Multimodal nocturnal seizure detection: do we need to adapt algorithms for children? *Epilepsia Open*. 2022;7(3):406–13.
41. Massé F, Bussel MV, Serteyn A, Arends J, Penders J. Miniaturized wireless ECG monitor for real-time detection of epileptic seizures. *ACM Trans Embedded Comput Syst*. 2013;12(4):1–21.
42. Munch Nielsen J, Zibrandtsen IC, Masulli P, Lykke Sørensen T, Andersen TS, Wesenberg KT. Towards a wearable multimodal seizure detection system in epilepsy: a pilot study. *Clin Neurophysiol*. 2022;136:40–8.
43. Tang J, El Atrache R, Yu S, Asif U, Jackson M, Roy S, et al. Seizure detection using wearable sensors and machine learning: setting a benchmark. *Epilepsia*. 2021;62(8):1807–19.
44. van Andel J, Ungureanu C, Arends J, Tan F, Van Dijk J, Petkov G, et al. Multimodal, automated detection of nocturnal motor seizures at home: is a reliable seizure detector feasible? *Epilepsia Open*. 2017;2(4):424–31.
45. Vandecasteele K, De Cooman T, Gu Y, Cleeren E, Claes K, Paesschen WV, et al. Automated epileptic seizure detection based on wearable ECG and PPG in a hospital environment. *Sensors*. 2017;17(10):1–12.
46. Zsom A, Lafrance WC, Blum AS, Li P, La W, Shaikh MA, et al. Ictal autonomic activity recorded via wearable-sensors plus machine learning can discriminate epileptic and psychogenic nonepileptic seizures. *Annu Int Conf IEEE Eng Med Biol Soc*. 2019;2019:3502–6. <https://doi.org/10.1109/EMBC.2019.8857552>
47. Al-Bakri AF, Villamar MF, Haddix C, Bensalem-Owen M, Sunderam S. Noninvasive seizure prediction using autonomic measurements in patients with refractory epilepsy. *Annu Int Conf IEEE Eng Med Biol Soc*. 2018;2018:2422–5.
48. Stirling RE, Grayden DB, D'Souza W, Cook MJ, Nurse E, Freestone DR, et al. Forecasting seizure likelihood with wearable technology. *Front Neurol*. 2021;12:704060.
49. Yamakawa T, Miyajima M, Fujiwara K, Kano M, Suzuki Y, Watanabe Y, et al. Wearable epileptic seizure prediction system with machine-learning-based anomaly detection of heart rate variability. *Sensors*. 2020;20(14):1–16.
50. Iman M, Arabnia HR, Rasheed K. A review of deep transfer learning and recent advancements. *Dent Tech*. 2023;11(2):40. <https://doi.org/10.3390/technologies11020040>
51. Fleming S, Thompson M, Stevens R, Heneghan C, Plüddemann A, Maconochie I, et al. Normal ranges of heart rate and respiratory rate in children from birth to 18 years of age: a systematic review of observational studies. *Lancet*. 2011;377(9770):1011–8.
52. van Elmpt WJ, Nijsen TM, Griep PA, Arends JB. A model of heart rate changes to detect seizures in severe epilepsy. *Seizure*. 2006;15(6):366–75.
53. Atwood AC, Drees CN. Seizure detection devices: five new things. *Neurol Clin Pract*. 2021;11(5):367–71.

54. Myers KA, Sivathamboo S, Perucca P. Heart rate variability measurement in epilepsy: how can we move from research to clinical practice? *Epilepsia*. 2018;59(12):2169–78.
55. Arbune AA, Jeppesen J, Conradsen I, Ryvlin P, Beniczky S. Perictal heart rate variability parameters as surrogate markers of seizure severity. *Epilepsia*. 2020;61(Suppl 1):S55–60.
56. Acharya UR, Hagiwara Y, Adeli H. Automated seizure prediction. *Epilepsy Behav*. 2018;88:251–61.
57. Bosl WJ, Leviton A, Loddenkemper T. Prediction of seizure recurrence. A note of caution. *Front Neurol*. 2021;12:675728.
58. Billeci L, Marino D, Insana L, Vatti G, Varanini M. Patient-specific seizure prediction based on heart rate variability and recurrence quantification analysis. *PLoS One*. 2018;13(9):e0204339.
59. Nasser M, Nurse E, Glasstetter M, Böttcher S, Gregg NM, Laks Nandakumar A, et al. Signal quality and patient experience with wearable devices for epilepsy management. *Epilepsia*. 2020;61(Suppl 1):S25–35.
60. Beniczky S, Wiebe S, Jeppesen J, Tatum WO, Brazdil M, Wang Y, et al. Automated seizure detection using wearable devices: a clinical practice guideline of the international league against epilepsy and the International Federation of Clinical Neurophysiology. *Clin Neurophysiol*. 2021;132(5):1173–84.
61. Bruno E, Böttcher S, Viana PF, Amengual-Gual M, Joseph B, Epitashvili N, et al. Wearable devices for seizure detection: practical experiences and recommendations from the wearables for epilepsy and research (WEAR) international study group. *Epilepsia*: John Wiley and Sons Inc; 2021. p. 2307–21.
62. Sivathamboo S, Nhu D, Piccenna L, Yang A, Antonic-Baker A, Vishwanath S, et al. Preferences and user experiences of wearable devices in epilepsy: a systematic review and mixed-methods synthesis. *Neurology*. 2022;99(13):e1380–92.

How to cite this article: Seth EA, Watterson J, Xie J, Arulsamy A, Md Yusof HH, Ngadimon IW, et al. Feasibility of cardiac-based seizure detection and prediction: A systematic review of non-invasive wearable sensor-based studies. *Epilepsia Open*. 2024;9:41–59. <https://doi.org/10.1002/epi4.12854>